




## **User Manual**

### **Oldelft Multiplane TE Probes**

Type numbers 19xT-, 19xV-, 19xZ-





---

## Table of Contents:

<b>1</b>	<b>Introduction</b> .....	<b>1</b>
1.1	Scope .....	1
1.2	Intended use of the multiplane TE Probe .....	1
1.3	Intention of this manual .....	2
<b>2</b>	<b>Preparing the multiplane TE Probe for Use</b> .....	<b>2</b>
2.1.	Visual examination .....	2
2.2.	Functional examination.....	2
2.3.	Safety precautions.....	3
2.4.	Preparation for Use .....	4
<b>3</b>	<b>Use of the multiplane TE Probe</b> .....	<b>5</b>
3.1.	Patient selection .....	5
3.2.	Introduction and withdrawal of the probe.....	5
3.3.	Articulation and scan plane rotation control.....	5
3.4.	Examination.....	7
<b>4</b>	<b>Cleaning and disinfecting the multiplane TE Probe</b> .....	<b>8</b>
4.1	Manual cleaning .....	8
4.2	Automated cleaning.....	8
4.3	Manual disinfection.....	9
4.4	Automated disinfection .....	9
4.5	Tristel Trio Wipes system .....	9
<b>5</b>	<b>Labeling</b> .....	<b>10</b>
<b>6</b>	<b>Storing the multiplane TE Probe</b> .....	<b>10</b>
<b>7</b>	<b>Specifications</b> .....	<b>11</b>
<b>Annex A.</b>	<b>Whom to Contact</b> .....	<b>12</b>




## 1 Introduction

This document is written as such that it is conform the EN 1041:2008 standard. Before using the equipment, carefully study this manual. Keep this manual at hand for quick reference. Reread periodically for warnings, safety precautions and maintenance requirements.


### 1.1 Scope

This manual is applicable for Oldelft Multiplane TE probes with part numbers in the 19xT-, 19xV and 19xZ- range. These probes have 32-64 elements and center frequency around 5 to 6 MHz. All are phased array transducers, mounted in a sealed tip at the end of a gastroscope. The probes are intended for imaging of the heart through the esophagus and the stomach. The array can be rotated 180° by a rotary knob on the handle of the probe. The articulation of the tip is controlled by a secondary rotary knob.



**figure 1. probe overview**

### 1.2 Intended use of the multiplane TE Probe

**WARNING**  **The multiplane TE Probe should be used only by a qualified physician who has received appropriate training in proper operation of the probe and in endoscopic techniques as dictated by current relevant medical practices.**

The physician conducting the examination must exercise sound medical judgement in the selection of patients for this probe and be skilled in interpreting the data obtained from the examination with the multiplane TE Probe.

The probe can be used to form images of the heart to detect abnormalities, to evaluate the velocity of the blood flowing in the heart and to obtain a color depiction of the velocities of the blood flowing in the heart.

### 1.3 Intention of this manual

This manual gives the user all information necessary to guarantee a safe and satisfactory use of the multiplane TE Probe. It also describes how to handle the probe during all stages : before, during and after the examination.

## 2 Preparing the multiplane TE Probe for Use

### 2.1. Visual examination

Visually examine and feel all portions of the probe before use, especially the gastroscope shaft and the flexible section at the distal end of the gastroscope. Perform the inspection of the flexible section both with the probe deflection straight and deflected. There should be no discontinuities, bumps, dents, holes, abrasions, bite-marks or any other evidence of wear or damage found.

The rigid tip at the distal end of the probe should be smooth and firmly attached to the gastroscope.

The cable and the connector that attach the probe to the ultrasound console should be free from evidence of damage.

### 2.2. Functional examination

Check the proper mechanical operation of the probe. Rotate the scan plane with the scan plane rotation controlling rotary knob on the handle (see **§Fout! Verwijzingsbron niet gevonden.**). Make sure that the scan plane rotation is smooth. Articulate the tip in both up and down direction using the articulation controlling rotary knob on the handle (see **§Fout! Verwijzingsbron niet gevonden.**). Make sure the articulation operates smoothly.

**Electrical  
Hazard**



Any evidence of damage indicates the probe cannot be used and should be returned to Oldelft for evaluation and repair.

## 2.3. Safety precautions

- a) The multiplane TE probe is a precision instrument, which must be handled with care. It may be damaged when dropped or abused. In particular keep sharp objects away from the acoustic window. Do not touch this window unnecessarily. Never exert force onto the window.
- b) Do not use a damaged probe. A damaged probe may cause an electrically hazardous condition when coupled to the human body. Inspect the probe for damage, cracks or bite-marks prior to each use.
- c) To avoid the risk of electric shock, this equipment must only be connected to a supply with protective earth.
- d) Only a qualified physician who has received appropriate training should perform examinations with the probe.
- e) The connector is not watertight, and should always be kept dry. The control handle, although spray-watertight, should not be immersed.
- f) This equipment contains no operator serviceable components. To prevent electric shock, do not remove any covers or panels.
- g) Never rub or spray the tip of the probe with an anesthetic agent.
- h) Avoid forceful manipulations and excessive force in using the probe, which could result in patient injury.
- i) Use a bite guard to prevent damage to the probe by the patients teeth.
- j) Only use water-based coupling gel.
- k) Withdraw the probe only with the friction deactivated and with the articulating section straight.
- l) Never manually articulate the distal tip of the probe directly; always use the articulation controlling rotary knobs.
- m) Do not kink, tightly coil, or apply excessive force on the probe cable or shaft. Insulation failure may result.
- n) The up/down and/or left/right articulation may after prolonged use develop an unwanted amount of free play. In that case, contact the service organization to re-adjust the articulation of the probe. This way, the risk of "buckling" or "U-turning" of the probe inside the esophagus is minimized.
- o) As with all diagnostic ultrasound procedures keep exposure levels and duration to the minimum necessary for the examination, especially in the Doppler modes.
- p) Clean and disinfect as described in this user manual. Do not use methods not described in this user manual like Ethanol, Iodine, Steam, Heat or Ethylene Oxide.
- q) Under normal conditions at full acoustic power the temperature of the tip does not exceed 43°C. To check that the temperature increase of the tip is within limits:

- |                                  |                |
|----------------------------------|----------------|
| <input type="radio"/>            | <b>Daily</b>   |
| <input type="radio"/>            | <b>Weekly</b>  |
| <input checked="" type="radio"/> | <b>Monthly</b> |
| <input type="radio"/>            | <b>Other</b>   |

- Connect the probe to the Ultrasound system.
- Adjust the Acoustic Power to the highest value possible.
- Select Color Doppler mode.
- Wait for 2 minutes.
- Feel at the distal end of the probe if there is a relevant temperature increase, which could be harmful for the patient.

## 2.4. Preparation for Use

Inspect the probe as described in §2.1.

Clean and disinfect the probe and the bite-guard as described in §4.

Make sure the system connector is plugged into the Ultrasound system and is locked by means of the locking handle on the connector or by means of a locking mechanism of the ultrasound system.

For patient protection a sterile, single-use, latex sheath can be used.

### CAUTION



Ensure that the sheath's sterile surface is maintained.

Devices containing latex may cause severe allergic reaction in latex sensitive individuals. Refer to FDA's March 29, 1991, "Medical Alert on Latex Products".

Place the bite guard on the probe so that after insertion of the probe the bite guard can easily be placed in the patient's mouth. The bite guard can also be placed in the patient's mouth before inserting the probe.

### WARNING



**The use of a bite-guard is mandatory. Failure to use the bite-guard may result in damage to the probe, which could result in a safety hazard. Damage to the probe due to biting is not covered by the probe's warranty.**

Apply a sufficient amount of water-soluble acoustic coupling gel on the probe acoustic window.

### CAUTION



Only use water-soluble acoustic coupling gel. Other coupling gels containing ingredients like ethanol, mineral oil, Iodine, lotions, lanolin, aloe vera or methyl or ethyl para benzoic acid can cause probe damage.

If used, place the latex sheath over the probe and gastroscope shaft up to but not covering the handle. Rub the tip carefully to ensure that all air bubbles have been removed from the probe's acoustic window area.

### CAUTION




Do not exert force on the window area.



### 3 Use of the multiplane TE Probe

#### 3.1. Patient selection

Although the multiplane TE probe can provide clinical data not available from other instruments, there are some considerations to be made in patient selection for safe use of the probe.

**CAUTION**  The ability of a patient to swallow or accommodate the probe should be considered.

Any history of gastro-esophageal diseases must be determined and considered as well as the possible effects of other therapies the patient is undergoing.

All gastro-esophageal abnormalities must be considered as well.

#### 3.2. Introduction and withdrawal of the probe


To protect the patient and the probe, unlock the knob friction on the knob for articulation when introducing or withdrawing the probe and make sure the articulating section is straight.

#### 3.3. Articulation and scan plane rotation control


The handle is designed for one-hand operation. Thumb, first and second fingers control the rotary knobs for articulation and scan plane rotation.

##### Tip articulation control

The largest rotary knob on the handle is for controlling the probe up/down tip articulation. When rotating the knob a click is felt at the central position, indicating the tip is in a straight position (no articulation).



**figure 2. tip articulation control**




**WARNING**




Check if the articulation of the tip is according specifications. If the articulation shows an unwanted amount of free play or exceeds the maximal articulation angles, do not use the probe. Contact the service organization to re-adjust the steering of the probe. In this way, the risk of "buckling" or "U-turning" of the articulating section in the esophagus is minimized.

**Brake operation**

The lower rotary knob has a freely moving mode and a mode where the knob has a friction. In the friction mode the movement of the articulation rotary knob is restrained. This can be used to hold the articulation in a certain position. The push button directly underneath the rotary knobs controls the friction function for the tip articulation. If the sliding bar shows blue<sup>1</sup> as depicted in the left picture below the brake is active.



**Brake active - blue<sup>1</sup>**



**Free moving articulation**

**figure 3. function of brake control button**


---

<sup>1</sup> The color can be different, depending on the probe type.

### Transducer scan plane rotation control


The smallest wheel on top is the scan plane rotation rotary knob.

The transducer scan plane can be rotated from 0° (transversal plane) via 90° (longitudinal plane) to 180° (transversal plane, left/right inverted). All planes in between can also be chosen. This allows for scanning a conical imaging volume.



### 3.4. Examination

The actual examination with the multiplane TE Probe is beyond the scope of this manual. There are many medical articles and books, which very thoroughly address this topic. There are however specific cautions that should be considered.

**CAUTION**  Long-term exposure to ultrasound should be minimized. Although there have been no confirmed adverse effects produced by diagnostic levels of ultrasound, unnecessary patient exposure to ultrasound energy should be avoided, especially in the Doppler mode.

Prolonged pressure on the esophagus by the tip of the probe may lead to a Pressure Necrosis phenomenon. Thus in monitoring applications the tip should be removed from the esophagus wall when you are not scanning by releasing it in the neutral position. If continuous monitoring is required the probe tip should often be re-positioned.

Whenever active scanning is not desired 'freeze' the image and release the articulation control.

Position the array such, that good acoustical contact is established. Before rotating the array, actuate the friction ('brake') on the up/down movement. If this is not done, rotating the array can mean loss of acoustical contact!

## 4 Cleaning and disinfecting the multiplane TE Probe

### Biological Hazard

Adequate cleaning and, if necessary disinfection are carried out to prevent disease transmission. It is the responsibility of the user to verify and maintain the effectiveness of the procedure used. A single-use, sterile disposable sheath for TE purposes can be used.

### CAUTION



Keep the control handle and system connector out of any cleaning or disinfection solutions. The control handle and cable may be cleaned with a damp cloth, but only the distal end of the probe up to the last marker on the shaft may be placed into a disinfection solution.

- |  |
|--|
| <ul style="list-style-type: none"><li><input type="radio"/> Daily</li><li><input type="radio"/> Weekly</li><li><input type="radio"/> Monthly</li><li><input checked="" type="radio"/> Other</li></ul> After each use |
|--|

For approved cleaning and disinfection agents, visit [www.oldelft.com](http://www.oldelft.com).

### 4.1 Manual cleaning

Sterile purified lukewarm water shall be used (maximum 26°C), unless otherwise stated by the cleaner manufacturer.

Perform the following steps, unless otherwise stated by the manufacturer of the selected cleaning agent:

1. Rinse the soiled probe (1 minute).
2. Soak the probe as prescribed by the manufacturer of the cleaning agent. Take care that only the distal end up to the last marker is immersed.
3. Gently scrub the probe using a soft-bristled brush (or wipes).
4. Rinse the probe for at least one 1 minute until there is no sign of blood or soil.
5. Dry the probe by means of a dry sterile wipe.

### CAUTION



Do not clean any portion of the probe with methanol or ethanol. Such substance can cause irreparable damage to the probe.

### 4.2 Automated cleaning

Sterile purified lukewarm water shall be used (maximum 26°C), unless otherwise stated by the cleaner manufacturer.

Before running automated cleaning, residual debris should be removed by immersing the used products in lukewarm water for 1 minute while removing coarse contamination with a fine, soft nylon brush or wipe. Take care that only the distal end up to the last marker is immersed.

Automated cleaning shall be performed in accordance with the instructions of the manufacturer of the cleaning agent and automated cleaner, using a validated cleaning cycle. Only the distal end of the TE probe shall be placed in the automated cleaner.

### 4.3 Manual disinfection

Sterile purified lukewarm water shall be used (maximum 26°C), unless otherwise stated by the manufacturer of the disinfection agent.

Perform the following steps, unless otherwise stated by the manufacturer of the disinfection agent:

1. Rinse the contaminated probe for at least one 1 minute.
2. Soak the probe for 12 minutes (or spray the probe if prescribed by the manufacturer of the disinfectant). Take care that only the distal end up to the last marker is immersed.
3. Gently wipe the probe.
4. Rinse the probe for at least one 1 minute.
5. Dry the probe by means of a dry sterile wipe.

#### CAUTION



The probe should never be left in the disinfection solution for more than 1 hour.

Do not forget to rinse the probe directly after disinfection.

### 4.4 Automated disinfection

Automated disinfection shall be performed in accordance with the instructions of the manufacturer of the disinfection agent and automated disinfectant / reprocessor, using a validated disinfection cycle for thermo-labile products.

The maximum allowable temperature is 50°C using peracetic acid based sanitizing solutions. Only the distal end of the TE probe shall be placed in the automated disinfectant / reprocessor.

### 4.5 Tristel Trio Wipes system

Cleaning and disinfection of the probe using the Tristel Trio Wipes system shall be performed in accordance with the instruction supplied with the wipes, except for the duration of the disinfection.

To obtain high level disinfection the duration of the disinfection step should be increased to 60s.

Take care that only the distal end up to the marker is exposed to the wipes.

#### WARNING



**Do not use other disinfection methods like Iodine, Steam, Heat or Ethylene Oxide.**

## 5 Labeling

The device should be labeled by our customer in such way that the labeling includes the following items. Preferably the symbols used are conform the EN 980. Furthermore, the labelling for products marketed in the USA should comply to UDI regulations.

- CE mark of the vendor;
- Reference to the Instruction for use;
- Serial number;
- Catalog number;
- Date Of Manufacture;
- Manufacturer;
- BF applied part symbol;
- Warning symbol.

## 6 Storing the multiplane TE Probe

The multiplane TE Probe is a sensitive and expensive instrument. Use great care when storing the probe. Make sure the probe is adequately cleaned and disinfected prior to storage.

The probe can be stored in its original case.

### WARNING



**Avoid damage to the probe by allowing nothing to protrude beyond the case when closing the lid.**

A wall mounted probe hanger can also be used for storage. Make sure the probe control handle, cable and connector are adequately supported. The shaft should hang down freely.

The storage environmental conditions must fall within the following range:

- Temperature between  $-25^{\circ}\text{C}$  and  $55^{\circ}\text{C}$ .
- Humidity between 5% and 95%, non-condensing.

## 7 Specifications

The probe has been designed to take into account all relevant provisions of the European Medical Device Directive 93/42/EEC as amended by Directive 2007/47/EC. The probe is categorized in Class IIa.

The probe is classified according to IEC 60601-1 as class I, type BF, defibrillator-proof.

	<b>Adult Multiplane</b>	<b>miniMulti</b>	<b>microMulti</b>
<b>Mechanical dimension</b>	<b>19xT-</b>	<b>19xV-</b>	<b>19xZ-</b>
» Shaft external diameter	11.2 mm	7.5 mm	5.2 mm
» Shaft length	100 cm	70 cm	80 cm
» Probe tip width	14.3 mm	10.9 mm	7.6 mm
» Probe tip height	10.6 mm	8.3 mm	5.6 mm
» Probe tip length	30.6 mm	27 mm	16 mm

<b>Transducer</b>			
» Center frequency	5.0 MHz	5.0 MHz	6.0 MHz
» Type	phased array		
» Number of elements	64	48	32
» Aperture	9 mm	7.25mm	5.2 mm
» Focus	50 mm	30 mm	40 mm

<b>Scan plane rotation</b>	minimum 180°
----------------------------	--------------

<b>Tip deflection</b>	
» Up	120° + 10°
» Down	90° + 10°

<b>Leakage current / dielectric strength</b>	Meet the requirements of IEC 60601
--	------------------------------------

<b>Biocompatibility</b>	All materials used in the patient applied part meet the requirements of ISO 10993
-------------------------	---

<b>Environmental requirements</b>	
» Operating temperature	0°C to 45°C
» Storage temperature	-25°C to 55°C
» Relative humidity	5 to 95%, non-condensing

## Annex A. Whom to Contact

### Manufacturer

Oldelft B.V.,  
Heertjeslaan 10,  
2629 JG Delft,  
The Netherlands.

Tel.: +31 15 2698 916

E-mail: [sales.ultrasound@oldelft.nl](mailto:sales.ultrasound@oldelft.nl)

Website: [www.oldelft.com](http://www.oldelft.com)