

# APPROVED DISINFECTANTS

Oldelft has tested the disinfectants listed below on chemical compatibility with the materials used in the patients applied part of our TEE probes. The patient applied part is the distal end of the probe up to the latest marker on the gastroscope shaft.

**Note:**

**The disinfectants are NOT tested on their biological effectiveness!**

For the biological effectiveness and the correct use of the cleaners and disinfectants, refer to the information supplied by the manufacturer of the disinfectants.

Disinfectant	Manufacturer	Active ingredient	Concentration	Approved for
Cidex	ASP J&J	Glutaraldehyde	2,55%	All TEE probes
Cidex OPA	ASP J&J	Ortho-phthalaldehyde	0,55%	All TEE probes
Gigasept AF	Schulke	Didecylmethylammoniumchloride Glycine, aminoalkyl derivs Tridecylpolyethyleneglycolether N-(3-Aminopropyl)-N-dodecylpropane-1,3-diamine	15% 6,9% 15-30% < 5%	All TEE probes
Gigasept FF (new)	Schulke	Dimethoxytetrahydrofuran Succindialdehyde Ethanol Methanol Alkylpolyethylenglykolpolypropyleneglycol-ether, 2-ethanol, 3,6-dioxa-1-dodecanol, 3,6-dioxadodecan-1-ol, DEGHE, diethylene glycol monohexyl ether, hexyl carbitol	3,2% 11,9% 5-15% 5-10% 1-5%	All TEE probes
Gigasept PAA concentrate	Schulke	Peracetic Acid Hydrogen peroxide Acetic Acid	5% 10-20% 10-20%	All TEE probes
Metricide	Metrex	Glutaraldehyde	2,60%	All TEE probes
Metricide 28	Metrex	Glutaraldehyde	2,50%	All TEE probes
Metricide Plus 30	Metrex	Glutaraldehyde	3,40%	All TEE probes
Metricide OPA Plus	Metrex	Ortho-phthalaldehyde	0,60%	All TEE probes
Perasafe	Dupont	Sodium Perborate	40-60%	All TEE probes
Rapicide AP HLD	Medivators	Peracetic Acid Hydrogen Peroxide Trisodium Phosphate Surfactant	5% 22% 4,3% 4%	microMulti- / miniMulti probe
Revital-Ox Resert XL HDL	Steris	Hydrogen Peroxide 2-Fluoroic Acid	1.4-2.3 ≤2.50	All TEE probes
Steranios 2%, 2% N.G., 2% E.C.S.	Laboratoires Anios	Glutaraldehyde	2,00%	All TEE probes
TD5	CS Medical	Glutaraldehyde	2,65%	All TEE probes
Tristel Trio Wipe System	Tristel	Propan-2-OL Polymeric Biguanide Hydrochloride 5-Chloro-2-Isotiazol-3-one 2-Methyl-2H-Isotiazol-3-O Chlorinedioxide Sodiumchlorite 100%	1-10% < 1% < 1% < 1% < 1%	All TEE probes

© 2011 Oldelft BV. All rights reserved. Copying of this document, and giving it to others and the use or communication of the contents thereof, are forbidden without express authority. Offenders are liable to the payment of damages. All rights are reserved in the event of the grant of a patent or the registration of a utility model or design. A printed copy is not a controlled document. Please see the PDM system for the approved version.

Rev.:	Date: 201809 27	Description: Remove disinfectants that are not available anymore.	Approval: RB
Designed by: K. de Vries		Category: Approved Chemicals	Sheet: 1 of 2
		Product: List of approved disinfectants	Doc. no. 250C002 01

# APPROVED DISINFECTANTS

Disinfectant	Supplier	Active ingredient	Concentration	Approved for
Wavicide 01	Medical Chemical Corp.	Glutaraldehyde	2,65%	All TEE probes

© 2011 Oldelft BV. All rights reserved. Copying of this document, and giving it to others and the use or communication of the contents thereof, are forbidden without express authority. Offenders are liable to the payment of damages. All rights are reserved in the event of the grant of a patent or the registration of a utility model or design. A printed copy is not a controlled document. Please see the PDM system for the approved version.

<i>Rev.:</i>	<i>Date:</i> 201809 27	<i>Description:</i> Remove disinfectants that are not available anymore.	<i>Approval:</i> RB
<i>Designed by:</i> K. de Vries		<i>Category:</i> Approved Chemicals	<i>Sheet:</i> 2 of 2
		<i>Product:</i> List of approved disinfectants	<i>Doc. no.</i> 250C002 01